BERHAMPORE GIRLS' COLLEGE BERHAMPORE, MURSHIDABAD

Tender Notice No. TN-/BGC/03/22

Date: 07.11.2022

Sealed quotations are invited from bonafide suppliers / manufacturers for supplying the following items. Quotations are to be given on or before 16.11.2022. All rates should be quoted including GST. Suppliers /manufacturers may also send quotations through mail (principal@berhamporegirlscollege.ac.in).

| Sl. | Item | Item Details | Quantity | Unit Price |
|-----|--|---|----------|---------------|
| 01 | Autoclave Size :- 12" Dia. X 20" Deep | Inner chamber:- 14SWG S. S. sheet 304 quality argon welded Lid and ring :- S.S. plate of 3/8" thick Lid :- Fly Nuts & Bolts with Rubber Gasket Pressure Gauge :- Yes Spring Loaded Safety Valve :- Yes Complete with S.S. basket :- Yes Line :- 220 / 230 V, 50 c/s. Single phase. Outer chamber:- M.S. sheet with powder coated finish Plate Thick :- 3/8" Pressure :- 5 to 20 p.s.i. | 01 | |
| | | All systems should be hydraulically tested upto 40 p.s.i. as a safety measurement. Steam Release Cock:- Yes Paddle Lifting Device:- Yes Automatic pressure cutout switch:- Yes | | |
| 02 | Laminar Air Flow Size :- 2X2X4 | Working Table: - S.S 304 Filter HEPA 0.3 micron of an efficiency of 99.97% down to 0.3 micron Pre filter: - Dry fibre washable type Standards International 209 E Blower: - Heavy duty Media: - Glass and made of pleats back in an anodized aluminium frame Noise level: - Low Velocity at the output: - HEPA is 90 ± 20 FPM Vibration: - Low Fluorescent Light: - Yes UV Light,: - Yes Static pressure Manometer: - Yes Gas cock: - Yes | 01 | |
| | | Front and side door :- ACRYLIC SHEET(Transparent) Front door :- Folding type | | |

Principal Berhampore Girls' College

Principal Bernampore Girls' College